

**Before using this device please read the following instructions completely and carefully. Correct application is vital to the proper functioning of this device.**

- 1)
  - A. Unfasten the Velcro straps at the forearm and wrist positions on the outside of the sling.
  - B. Insert arm into the sling, resting the cushion on your hip. Place your arm as far back as possible in the sling and the thumb strap between the thumb and first finger.
  - C. Reattach both the forearm and wrist straps on top of the sling to secure the arm.
- 2) Align cushion on the injured side at waist level with the ball (front) away from your body. Place the pillow so the line on top of the pillow is parallel to the front of your body. (If you stand at a table, the line should be parallel to the table.)
- 3)
  - A. Bring waist strap around from the back of cushion around your waist, inserting into and attaching to the front buckle.
  - B. Adjust waist strap for proper fit.
- 4) Unfasten the shoulder strap buckle at the front of the cushion. Using your free arm, reach behind your body and slip your arm through the shoulder strap. This motion will be similar to putting on a backpack, with the wide strap on the clavicular area and the thin pad in the underarm area.
- 5) Reattach the shoulder strap to buckle at front of cushion.
- 6) Adjust all straps for adequate stabilization of injured shoulder.

### Protocol:

- 1) Read insert before use. Follow manufacturer's instructions, hospital protocol, and state or federal laws.
- 2) Follow physician's orders for appropriateness, duration of application and frequency of use.
- 3) Train all staff on proper application.
- 4) Do not alter DJO Global UltraSling III ER in any way or transfer products from one facility to another.
- 5) Do not expose to open flame.

In addition to the UltraSling III ER DJO Global offers a complete line of shoulder immobilization products.

### INTENDED USE/INDICATIONS:

For protocols involving immobilization in external rotation: Glenohumeral dislocation/subluxation, Capsular Shifts, Use after posterior shoulder stabilizations, Postoperative Bankart procedures, Soft tissue strains/repairs, Post operative positioning after release of severe anterior capsule contracture.

### CONTRAINDICATIONS:

None.

### WARNINGS AND PRECAUTIONS:

This product is to be used under the supervision of a medical professional. This device is not intended for unsupervised public use. If you experience any pain, swelling, sensation changes, or any unusual reactions while using this product, consult your medical professional immediately.


### WARRANTY:

DJO, LLC will repair or replace all or part of the unit and its accessories for material or workmanship defects for a period of six months from the date of sale.

### CLEANING INSTRUCTIONS:

Hand wash with mild soap in lukewarm water, rinse thoroughly. AIR DRY.

Note: If not rinsed thoroughly, residual soap may cause irritation and deteriorate material.



### UltraSling III ER

Maintains the shoulder and arm in an externally rotated position. The new moisture wicking mesh fabric offers greater comfort, particularly in longer term use. 15° and 30° versions are available to suit your protocol. The De-Rotation Strap keeps product in place. Quick release buckles make it easy to take on and off for exercise/therapy.

Part Number		Size
11-1343-X	15°, Right	S - XL
11-1344-X	15°, Left	S - XL
11-1345-X	30°, Right	S - XL
11-1346-X	30°, Left	S - XL
11-9100	15° Wedge,(each)	S - XL
11-9101	15° Wedge, (5-pack)	S - XL

*Measure below elbow crease to base of index finger, fits right or left arm. x = Size*

x = 2 = S	Up to 11"	(28cm)
x = 3 = M	11" - 13"	(28 - 33cm)
x = 4 = L	13" - 15"	(33 - 38cm)
x = 5 = XL	15"+	(38cm+)

For single patient use only. (RX Product) Latex free

For more information, please contact your DonJoy Territory Manager or Authorized Distributor



DJO, LLC | A DJO Global Company

T 800.336.6569 | D 760.727.1280 | F 800.936.6569

1430 Decision Street | Vista, CA 92081-8553 | U.S.A.

[DJOglobal.com](http://DJOglobal.com)

*Together in Motion™*